

This work is on a Creative Commons Attribution 4.0 International (CC BY 4.0) license, <https://creativecommons.org/licenses/by/4.0/>. Access to this work was provided by the University of Maryland, Baltimore County (UMBC) ScholarWorks@UMBC digital repository on the Maryland Shared Open Access (MD-SOAR) platform.

Please provide feedback

Please support the ScholarWorks@UMBC repository by emailing [scholarworks-group@umbc.edu](mailto:scholarworks-group@umbc.edu) and telling us what having access to this work means to you and why it's important to you. Thank you.



# CAFE: a new, improved nonresonant laser-induced fluorescence instrument for airborne in situ measurement of formaldehyde

Jason M. St. Clair<sup>1,2</sup>, Andrew K. Swanson<sup>1,3</sup>, Steven A. Bailey<sup>1</sup>, and Thomas F. Hanisco<sup>1</sup>

<sup>1</sup>Atmospheric Chemistry and Dynamics Laboratory, NASA Goddard Space Flight Center, Greenbelt, MD, 20771, USA

<sup>2</sup>Joint Center for Earth Systems Technology, University of Maryland Baltimore County, Baltimore, MD, 21228, USA

<sup>3</sup>Goddard Earth Sciences Technology and Research, Universities Space Research Association, Columbia, MD, 21046, USA

**Correspondence:** Jason M. St. Clair (jason.m.stclair@nasa.gov)

Received: 19 April 2019 – Discussion started: 20 May 2019

Revised: 29 July 2019 – Accepted: 6 August 2019 – Published: 30 August 2019

**Abstract.** NASA Compact Airborne Formaldehyde Experiment (CAFE) is a nonresonant laser-induced fluorescence instrument for airborne in situ measurement of formaldehyde (HCHO). The instrument is described here with highlighted improvements from the predecessor instrument, COmpact Formaldehyde Fluorescence Experiment (COFFEE). CAFE uses a 480 mW, 80 kHz laser at 355 nm to excite HCHO and detects the resulting fluorescence in the 420–550 nm range. The fluorescence is acquired at 5 ns resolution for 500 ns and the unique time profile of the HCHO fluorescence provides measurement selectivity. CAFE achieves a  $1\sigma$  precision of 160 pptv (1 s) and 90 pptv (10 s) for  $[\text{HCHO}] = 0$  pptv. The accuracy of CAFE, using its curve-fitting data processing, is estimated as  $\pm 20\%$  of  $[\text{HCHO}] + 100$  pptv. CAFE has successfully flown on multiple aircraft platforms and is particularly well-suited to high-altitude research aircraft or small air quality research aircraft where high sensitivity is required but operator interaction and instrument payload is limited.

## 1 Introduction

Formaldehyde (HCHO) in Earth's atmosphere is primarily produced during the oxidation of volatile organic compounds (VOCs) including methane and isoprene, with minor direct emissions from biomass burning and fossil fuel combustion (Fortems-Cheiney et al., 2012). The atmospheric lifetime of HCHO is a few hours against photolysis and reaction with OH. HCHO is thus sensitive to local variability in both VOC emissions/transport and oxidation capacity. HCHO measurements are consequently valuable for deriving continental-

scale isoprene emission fluxes (Palmer et al., 2003), remote-atmosphere maps of column OH (Wolfe et al., 2019), and regional-/continental-scale maps of organic aerosol (Liao et al., 2019). Deep convective transport of oxygenated VOCs (OVOCs) such as HCHO (e.g., Fried et al., 2016), or transport of OVOC precursors (e.g., Apel et al., 2012), can account for a significant portion of the upper tropospheric (UT)  $\text{HO}_x$  budget (e.g., Ren et al., 2008) and thereby impact UT  $\text{O}_3$  levels (e.g., Crawford et al., 1999). HCHO measurements in the UT are necessary to close the  $\text{HO}_x$  budget (Jaeglé et al., 2001) and can also be used in the UT and lower stratosphere (LS) as a qualitative tracer of recent convective activity. Monitoring HCHO is also important for public health: the EPA classifies HCHO as a hazardous air pollutant (HAP) and it is estimated to pose the highest cancer risk of all the HAP species (Strum and Scheffe, 2016).

A number of measurement techniques have been applied to in situ atmospheric measurements of HCHO and are reviewed elsewhere (Fried et al., 2008; Hak et al., 2005; Kaiser et al., 2014; Shutter et al., 2019). The most common techniques are absorption spectroscopy (e.g., Richter et al., 2015; Washenfelder et al., 2016), wet chemistry (e.g., Aiello and McLaren, 2009), and laser-induced fluorescence (LIF) (e.g., Cazorla et al., 2015; Hottle et al., 2009). The different techniques offer trade-offs between cost, instrument size, sampling time resolution, and sensitivity. A new approach, non-resonant LIF (NR-LIF), was recently developed to provide a more compact, lower-cost instrument with simpler operation than typical HCHO LIF instruments (St. Clair et al., 2017). The NR-LIF approach trades the wavelength-tunable excitation laser traditionally used in HCHO LIF for a fixed-

wavelength 355 nm laser that excites multiple HCHO absorption features. The broad, fixed-wavelength laser cannot tune on and off a single absorption feature to uniquely detect HCHO. Measurement selectivity is provided by carefully selected fluorescence optical filters and rapid data acquisition that enables us to resolve the time profile of the HCHO fluorescence signal.

The NR-LIF HCHO instrument described in St. Clair et al. (2017), Compact Formaldehyde Fluorescence Experiment (COFFEE), was specifically designed for operation on the Alpha Jet stationed at NASA Ames Research Center and is not easily integrated into other aircraft platforms. The Compact Airborne Formaldehyde Experiment (CAFE) instrument described here uses the same general measurement approach as COFFEE but with updated components and design changes for operation under the demanding environmental conditions of the NASA ER-2 aircraft. The compact instrument design combined with reliable, commercial subsystems makes CAFE well-suited for scientific projects on high-altitude aircraft, where high sensitivity is required but size and weight are limited, as well as tropospheric aircraft platforms with limited payload. In this paper we present the instrumental details of CAFE, including modifications from the original COFFEE design that stem from an improved understanding of the HCHO NR-LIF technique.

## 2 Instrument description


The CAFE instrument is enclosed by a 42 cm × 53 cm × 30 cm ( $L \times W \times H$ ) chassis (Fig. S1 in the Supplement) with all plumbing and electrical connections on one side of the instrument. The main subsystems of the instrument are shown in Fig. 1 with the chassis lid removed and are described in more detail in the following sections.

### 2.1 Laser

The light source in CAFE is a Spectra-Physics Explorer One 355-1W (Fig. 1, item A) that provides 480 mW of pulsed 355 nm light. The laser is operated at 80 kHz (adjustable 0–300 kHz) to reduce the peak power and minimize risk of mirror damage. Pulses are  $\sim 10$  ns wide (full width at half maximum) and the laser bandwidth is approximately 1 nm. Laser control is performed by a RIO data acquisition system (Sect. 2.5) through RS232 communication.

### 2.2 Optical system


Figure 2 illustrates the details of the CAFE optical plate (Fig. 1, item B). The laser beam is directed into the optical cell by three dielectric (Acton Optics) steering mirrors (Fig. 2, item A) with an antireflection (AR) coating. A collimating lens ( $F = 100$  mm) and half wave plate (AR coated, Optisource) are positioned in between mirrors 2 and 3 (to-



**Figure 1.** A top view of the CAFE instrument, with major components labeled: (A) laser, (B) optical plate, (C) pressure controller, (D) pump, and (E) RIO data acquisition system.


gether Fig. 2, item B). AR-coated fused silica (CVI) cell windows (Fig. 2, item C) provide the vacuum seal for the optical entrance and exit of the optical cell. Optical baffles (Fig. 2, item D) along the entrance and exit “baffle arms” of the optical cell significantly reduce the stray light reaching the detection volume. Baffle opening diameter increases from entrance to exit over the seven baffles with three sizes (2.5, 3, 3.5 mm) to accommodate any beam divergence. The baffles closest to the detection volume are designed in-house and commercially coated (Acktar, vacuum black) while the other five baffles are commercially available spatial filter apertures (National Aperture). Cylindrical spacers between the baffles are coated (Acktar, magic black) though background counts are minimally sensitive to the spacer surface for spacers not adjacent to the detection volume. The inner surface of the detection cell is anodized black to reduce stray light. To achieve a more compact optical layout, the ends of the baffle arms are at  $90^\circ$  to the rest of the arm. The  $90^\circ$  turns are provided by two fixed steering mirrors (Fig. 2, item A). After exiting the optical cell, the laser beam passes through a beam sampler (Thorlabs) (Fig. 2, item E), which directs a few percent of the beam to a power meter (Fig. 2, item F) comprised of a ground glass diffuser (Thorlabs DGUV 10-600), an absorption filter (Thorlabs FGUV11), and an amplified photodiode (OSI 555-UV). The majority of the laser beam passes through the beam sampler and into a beam dump.

Sample air flows through the detection cell perpendicular to the laser beam path and detection axes. Aspheric lenses (NA = 0.66, AR-coated, Edmund Optics) on opposite sides of the detection cell (Fig. 2, item G) image the intersection of the sample air flow and the laser path to collect the HCHO fluorescence ( $\sim 355$ –550 nm). For each detection arm, the collected light is turned  $90^\circ$  by a mirror (right angle prism



**Figure 2.** CAFE instrument top view, with dust covers removed and a cutaway of the detection cell to show its optical components: (A) steering mirrors, (B) half-wave plate and collimating lens, (C) cell windows, (D) laser baffles, (E) beam sampler, (F) laser power monitor, (G) aspheric lens, (H) prism dielectric mirror, (I) optical filters, (J) lens, and (K) photomultiplier tube.

dielectric, Thorlabs) (Fig. 2, item H) and directed toward the optical filters (Fig. 2, item I), a lens ( $F = 75$  mm, AR-coated, Thorlabs) (Fig. 2, item J), and the photomultiplier tube (PMT) (Hamamatsu H7360-02 MOD) (Fig. 2, item K). The optical filters for the two detection axes differ from each other and from the filters described in St. Clair et al. (2017). The filters were selected to transmit the maximum amount of HCHO fluorescence while blocking the laser wavelength and Raman scatter from  $N_2$ ,  $O_2$ , and  $H_2O$ , as shown for Axis 1 in Fig. 3a and Axis 2 in Fig. 3b. The primary motivation for the change in filters from St. Clair et al. (2017) was the realization that  $H_2O$  Raman signal, at high  $[H_2O]$ , adversely affected data processing as discussed in the Supplement (Fig. S2). Replacing some of the long-pass filters with red-shifted alternatives reduces the instrument sensitivity but improves accuracy at higher ambient water vapor. Axis 1, proceeding toward the PMT, uses an AR-coated 370 nm long-pass absorption filter (Hoya Candeo Optonics), a 405 nm long-pass interference filter (Semrock), and a 420 nm long-pass absorption filter (Thorlabs). Axis 2, proceeding toward the PMT, has an AR-coated 370 nm long-pass absorption filter (Hoya Candeo Optonics), a custom 11-band band pass interference filter (Semrock), a 405 nm long-pass interference filter (Semrock), and a 420 nm long-pass absorption filter (Thorlabs). As of April 2018, COFFEE also uses these fluorescence filters. Axis 2, with the 11-band band pass filter, is  $\sim 3\times$  less sensitive than Axis 1. The 11-band filter, however, provides greater confidence in the nonresonant LIF measurement selectivity. With more ambient obser-




**Figure 3.** The combined optical filter transmission spectrum for (a) detection Axis 1 and (b) detection Axis 2. HCHO fluorescence and  $N_2$ ,  $O_2$ , and  $H_2O$  Raman spectra are shown with arbitrary units. The Raman spectra are for the first Stokes Q branch only and are intended to give a general sense of the transition wavelengths. Filters transmit approximately 19 % and 34 % of the fluorescence for Axis 1 and Axis 2, respectively.

vations, the Axis 1 filters may prove to be sufficiently selective to be used on both axes.

### 2.3 Air sampling

The details of the inlet and sample tubing external to the instrument are dependent on the aircraft platform for a given experiment. CAFE has used previously existing inlets for various aircraft platforms, e.g., the inlet described in Cazorla et al. (2015) for the NASA DC-8. A new inlet was designed in 2016 for use on the high-altitude NASA ER-2 aircraft. The inlet body consists of a nacelle mounted to the end of an air foil that attaches to the ER-2 superpod midbody via an adapter plate (Fig. 4). The center of the nacelle is held 10.1 cm from the adapter plate to extend sampling beyond the boundary layer of the superpod. Ram pressure is generated from the tapered shape of the nacelle and allows a



**Figure 4.** A cross-sectional view of the CAFE inlet used on the NASA ER-2. Ram pressure is created by ambient air entering a nacelle mounted on the end of an air foil. Sample air is drawn toward the instrument through a heated Silcosteel tube that protrudes into the nacelle perpendicular to the ambient air flow. Internal to the pod, 1.3 m of tubing delivers the sample flow to the instrument and an equal amount returns the exhaust flow to the inlet where it is directed out the back of the air foil.

higher instrument operating pressure than possible with ambient pressure sampling. Sample air is pulled into the instrument through a heated 6.35 mm OD Silcosteel tube that protrudes into the nacelle perpendicular to the ambient air flow. The tube is chamfered  $15^\circ$  away from the direction of flight to reject particles larger than  $\sim 1 \mu\text{m}$ . A total of 28 W of heat is provided by a cartridge heater (SunRod) mounted in a copper block that clamps around the sample tube, and is controlled to  $30^\circ\text{C}$  by a proportional controller (Minco CT335) with protection from thermal run-away using a bimetallic hermetic thermostat.

Inside the pod, 1.3 m of tubing (fluorinated ethylene propylene (FEP) or stainless steel) carries the sample flow from the inlet to the instrument. A particle filter (Balston 9922-05-DQ) in line with the tubing mitigates aerosol-related artifacts (St. Clair et al., 2017). Sample air enters CAFE through a Swagelok VCO bulkhead fitting and 7 cm of FEP tubing before reaching a pressure controller (Fig. 1, item C). The pressure controller consists of a custom Ni-plated valve block and an actuator (iQ Valve) with MKS control electronics. A length of 22 cm of FEP tubing connects the pressure controller to the detection cell. The final section of tubing is covered in an opaque black shrink-wrap to reduce background counts from ambient light while the chassis cover is off for alignment. All metal surfaces before the detection cell are coated with a fluorocarbon (FluoroPel, Cytonix) to reduce HCHO surface interactions and improve instrument time response. For tropospheric measurements, detection cell pressure is maintained at 10.7 kPa by the pres-

sure controller and mass flow through the instrument is typically  $\sim 2.3$  standard  $\text{L min}^{-1}$  (sLm). For stratospheric measurements, the cell pressure is necessarily reduced to 4 kPa to maintain control at high altitudes. Air passes through the detection cell and along the bottom of the optical plate where it proceeds to the pump (Vacuumbrand MD-1, Fig. 1 item D), first passing through a  $\sim 70 \text{ cm}^3$  ballast volume that muffles the pressure fluctuations from the pump. A fraction of the airflow leaves the detection cell via the baffle arms to flush that volume and rejoins the main flow after the ballast volume. Pump exhaust is plumbed from the instrument to the inlet where it is directed out the back of the air foil (Fig. 4).

## 2.4 Thermal management

Temperature control, to varying degrees, is necessary for some of the instrument subsystems. Adequate heat removal is required for proper laser performance: two thermoelectric cooler (TEC) elements (TE Technology) provide unidirectional thermal control to cool the laser mounting plate and maintain temperature to  $\leq 30^\circ\text{C}$ . TECs are thermally attached to the optical plate with additional heat dissipation provided by heat sinks mounted to the underside of the optical plate. Air is drawn up through the bottom of the instrument chassis by two fans ( $22 \text{ L s}^{-1}$  each) that blow directly onto the heat sinks. The fans are controlled by a  $30/35^\circ\text{C}$  off/on controller (NuWave Technologies) to avoid cooling at low temperatures. The pump is a significant heat source and so the four fans that exhaust air out of the chassis (visible in Fig. S1) pull air across the pump to prevent its heat from adversely affecting the laser and other subsystems. The exhaust fans are also controlled to turn off below  $30^\circ\text{C}$ .

The pump operates best when the ambient temperature is  $10\text{--}40^\circ\text{C}$ . To maintain pump function in the unheated pods of high-altitude aircraft, 31 W of cartridge heaters is installed on the pump. Cartridge heaters are also used to heat the pressure controller (24 W) and detection cell (24 W) to  $30^\circ\text{C}$ . Polyimide foil heaters provide 70 W of heat to the optical plate, also maintained at  $30^\circ\text{C}$ . All heaters are controlled by proportional controllers (Minco CT335) and protected from thermal run-away using thermostats.

## 2.5 Data acquisition

CAFE utilizes a National Instruments CompactRIO (RIO) data acquisition system (Fig. 1, item E) that runs a real-time operating system. Signal from each PMT is counted with 5 ns resolution by two channels of a NI 9402 high-speed digital I/O module, which is triggered by a transistor–transistor logic (TTL) pulse from the laser. The signals from the PMTs are delayed by 50 ns with a passive delay circuit (Data Delay Devices, 1515 series) so that the PMT signals arrive after the laser TTL pulse. The digital I/O module, along with analog NI 9205 (input) and NI 9264 (output) modules, interfaces with a main processor module and a backplane with a




field-programmable gate array (FPGA). Data are recorded at 10 Hz for non-gated (continuous) and time-gated counts and at 1 Hz for all other data including instrument diagnostics and time-resolved counts. The 10 Hz data are used mainly as an alignment metric. The 1 Hz time-resolved count data consist of a 500 ns time profile of counts, resolved to 100 discrete 5 ns bins, immediately following the laser TTL pulse. HCHO mixing ratio data are generated from the 1 Hz time-resolved count data.

## 2.6 Data processing

The data processing approach for CAFE is almost identical to the methods described in St. Clair et al. (2017) for COFFEE with minor changes related to laser pulse timing unique to each instrument. Briefly, data from each detection axis are processed by performing three sequential fits to the 1 Hz, 5 ns binned data (0–500 ns from trigger). The first fit is a single-parameter fit to remove the slowly decaying “long-lived” background signal from the time profile and give a zero baseline at a high bin number where HCHO does not contribute to the signal. After the long-lived signal is removed, the 0–500 ns time profile is fit using a two-parameter nonlinear least-squares optimization as shown in Fig. 5 for Axis 2. The two parameters of the fit are scaling factors to the profiles that represent the HCHO (dashed red) and non-HCHO “air” (dashed blue) portions of the time profile, which are determined by laboratory calibration and a preflight zero run. While the representative profiles, or exemplars, span the full 500 ns acquisition period, the fit (black line) is performed only from bin 11 to bin 56 (50–280 ns) in a trade-off between HCHO data precision and fit quality. The third step is a single-parameter nonlinear least-squares fit with only the HCHO portion of the fit varying and the non-HCHO portion fixed to a smoothed version of the two-parameter fit output, improving the precision of the HCHO data compared to the output of the two-parameter fit. The calibration scalar is then applied to the single-parameter fit output to yield HCHO in parts per billion by volume. Mixing ratio data from the two axes are averaged together to provide the final data product.

Data can be processed using delayed-gate counts at 1 or 10 Hz to give higher-precision HCHO data (see Sect. 3.2) that are potentially more susceptible to measurement artifacts. Delayed-gate count data sum the PMT counts over the same time window as the exemplar fit (bin 11 to bin 56, 50–280 ns), keeping the background signal low by excluding the prompt signal from scatter. HCHO data in parts per billion by volume are generated by subtracting the background counts determined from a preflight zero run and applying the calibration factor. More details on the fits, the “exemplar” time profiles, and processing with gated counts are available in St. Clair et al. (2017).



**Figure 5.** An example of a 1 Hz, 5 ns binned time profile (cyan circles) for  $\sim 10$  ppbv HCHO is fit using a two-parameter nonlinear least-squares approach. The two components of the fit are the scaled air (dashed blue line) and HCHO (dashed red line) contributions to the time profile. The fit, shown in black, is performed over a subset of the time profile.

## 2.7 Summary of updates from COFFEE

CAFE applies the same measurement technique and uses many of the same components as COFFEE, but also contains a number of updates. The updates are summarized here to aid the reader in understanding the differences between the two instruments. CAFE uses a new version of the Explorer One laser (Sect. 2.1) operated at twice the repetition rate (80 kHz), yielding a 60 % increase in average laser power. The fluorescence filters (Sect. 2.1) were changed to block Raman scattering signal from water vapor that caused a fitting bias in data processing. The additional filters, which decrease the instrument sensitivity to HCHO, are necessary for tropospheric observations but not for UT–LS observations where water vapor is low. The shape of CAFE’s chassis (Fig. S1) is more compact and easier to integrate into most aircraft platforms; the unusual shape of COFFEE is due to the available volume in the Alpha Jet pod. The compact CAFE chassis is made possible by including two additional turning mirrors at 90° bends in the baffle arms. A ballast volume (Sect. 2.1) between the detection cell and the pump was added to muffle pressure fluctuations from the diaphragm pump and prevent them from propagating into the detection cell. Heaters were added to the pump (Sect. 2.4) to maintain its optimal operation temperature during high-altitude flights.


### 3 Performance

#### 3.1 Calibration and sensitivity

CAFE calibration is performed in the same manner as for the In Situ Airborne Formaldehyde (ISAF) and COFFEE instruments. Flow ( $0\text{--}50\text{ standard cm}^3\text{ min}^{-1}$ , sccm) from a  $\sim 500\text{ ppbv}$  HCHO in  $\text{N}_2$  gas cylinder is added to a main flow ( $\sim 4\text{ sLm}$ ) of ultrahigh-purity air to give HCHO of typically  $\sim 0\text{--}10\text{ ppbv}$ . The  $\sim 4\text{ sLm}$  flow is greater than the instrument sample flow ( $\sim 2.3\text{ sLm}$ ) to speed up the time response of the calibration set-up, with the excess flow exhausting to the room through a “T” at the instrument chassis. The HCHO mixing ratio is stepped through three to five different flows with  $>45\text{ min}$  at each level to provide confidence that the calibration set-up has equilibrated. The mixing ratio of the HCHO calibration cylinders is verified before and after each aircraft mission using infrared (IR) absorption as described in Cazorla et al. (2015). Uncertainty in the gas standard combines with uncertainty in the calibration system flow measurements to give a calibration uncertainty of  $\pm 10\%$ . For the purposes of comparison with other LIF instruments, the ungated count sensitivities are  $1.3\text{ counts s}^{-1}\text{ mW}^{-1}\text{ ppbv}^{-1}$  (Axis 1) and  $0.39\text{ counts s}^{-1}\text{ mW}^{-1}\text{ ppbv}^{-1}$  (Axis 2). The gated sensitivities are  $0.833\text{ counts s}^{-1}\text{ mW}^{-1}\text{ ppbv}^{-1}$  (Axis 1) and  $0.24\text{ counts s}^{-1}\text{ mW}^{-1}\text{ ppbv}^{-1}$  (Axis 2). Sensitivity differences from COFFEE are due to the combined effects of multiple factors: the changes in optical filters used, differences in individual PMT sensitivities, and widening the time gate. A discussion of ISAF sensitivities is included in St. Clair et al. (2017).

#### 3.2 Precision and measurement uncertainty


The main source of noise for CAFE HCHO is Raman and Rayleigh scattering of the laser beam. Chamber scatter also significantly contributes to the noise and varies in magnitude with the alignment of the laser through the optical cell. Figure 6 shows the standard deviation of HCHO as a function of HCHO mixing ratio for data processed using exemplar fitting (red symbols) and gated counts (blue symbols) from multiple laboratory experiments using ultrahigh-purity (UHP) air. At  $[\text{HCHO}] = 0\text{ ppbv}$ , the average standard deviation for the exemplar-fitted data was  $\pm 152\text{ pptv}$  for 1 s and  $\pm 87\text{ pptv}$  for 10 s data. For gated count data, the standard deviation was lower:  $\pm 100\text{ pptv}$  for 1 s and  $\pm 37\text{ pptv}$  for 10 s data. These values are somewhat higher than reported in St. Clair et al. (2017) for COFFEE despite CAFE’s higher laser power due to the changes in optical filters and the increased delay from the laser pulse to the start of the fit window in CAFE data processing. Using the same fitting window and gate delay but with the fluorescence filters from St. Clair et al. (2017), the average standard deviation at  $[\text{HCHO}] = 0\text{ ppbv}$  for a laboratory experiment was  $\pm 106\text{ pptv}$  for 1 s and  $\pm 62\text{ pptv}$  for 10 s data



**Figure 6.** Standard deviation as a function of HCHO for two data processing techniques: exemplar fitting for 1 s (red circle) and 10 s (red x) data, and gated counts for 1 s (blue circle) and 10 s (blue x) data.

(exemplar-fitted) and  $\pm 55\text{ pptv}$  for 1 s and  $\pm 18\text{ pptv}$  for 10 s data (gated data). The precision can be improved by shortening the delay between the laser pulse and the start of the fit window, thereby increasing the number of fluorescence data used in the fit. With the  $\text{H}_2\text{O}$  Raman signal effectively blocked, shortening the delay may now be appropriate but further investigation is necessary. For UT-LS measurements of HCHO, the requisite precision exceeds what is possible using exemplar fitting with the current fluorescence filters. Blocking signal from  $\text{H}_2\text{O}$  Raman, however, is not necessary at the low specific humidity of the UT-LS and the water-blocking long-pass filters will not be used for high-altitude projects, thereby increasing the fluorescence signal. The gated count data also provide an improvement in precision over the exemplar fits, as shown in Fig. 6, and will be available for higher signal-to-noise measurements in the UT-LS.

Figure 7 shows the Allan–Werle deviation as a function of averaging time for two-parameter fit data with  $[\text{HCHO}] = 6\text{ ppbv}$  to illustrate how precision improves with increased data averaging. HCHO data precision improves with averaging up to  $\sim 200\text{ s}$ , after which additional time averaging has no benefit. A linear fit to the data with  $\tau \leq 256\text{ s}$  gives a slope ( $-0.58$ ) similar to the value expected from white noise (0.5), suggesting that white noise dominates on those timescales. The overall HCHO measurement uncertainty is a combination of the calibration uncertainty ( $\pm 10\%$ ) and the less easily quantified uncertainties from unidentified measurement interferences and fit biases. To account for both the known and unknown uncertainties, the overall measurement uncertainty is conservatively estimated to be  $\pm (20\% \text{ of } [\text{HCHO}] + 100\text{ pptv})$ . For specific projects that require higher accuracy, the instrument must



**Figure 7.** Normalized Allan–Werle deviation versus averaging time for  $[\text{HCHO}] = 6$  ppbv. HCHO precision improves with averaging up to  $\sim 200$  s. A linear fit (red dashed line) to the data up to 256 s gives a slope ( $-0.58$ ) similar to the value expected from white noise ( $0.5$ ).

be configured to reduce the 100 pptv term in the uncertainty. The instrument can be zeroed in-flight, either by using a scrubber such as 2,4-dinitrophenylhydrazine (DNPH) or Drierite + molecular sieve, or by flying a zero-air canister. The addition of a zeroing system will likely impact the instrument time response.


### 3.3 Time response

For airborne in situ observations, the effective spatial resolution is limited by both the data acquisition rate and the time response of the instrument. To understand the time response of CAFE, pulses of 1.3 ppmv HCHO gas were introduced into a high flow of UHP air by a rapidly switching valve (The Lee Company, IEP series) and CAFE subsampled this flow. The decay of the 10 Hz HCHO signal after the pulses was fit using an exponential function, as shown in Fig. 8, and the measured instrument time response is  $158 \pm 10$  ( $1\sigma$ ) ms. An additional set of measurements was conducted for CAFE with the particle filter in line with the sample flow to verify that the filter does not adversely affect the instrument time response. With the filter in line, the  $e$ -folding time (161 ms) was nearly identical to the response without the filter.

## 4 Field deployment

### 4.1 Aircraft integration

CAFE has successfully flown on the NASA DC-8, ER-2, and C-23 Sherpa aircraft, the Hanseo University King Air during KORUS-AQ, and the University of Maryland Cessna 402B. The rectangular box shape of the chassis (Fig. S1) and internal pump make CAFE a simple instrument to integrate.



**Figure 8.** Pulses of HCHO were fit to provide the  $e$ -folding flush time ( $158 \pm 10$  ms,  $1\sigma$ ) of the instrument.


Fasteners attach CAFE to structural angles from below, using mounting holes with blind locking nut plates on the bottom of the instrument. For tropospheric aircraft projects, the angles are riveted to non-pivoting cabinet slides (General Devices) to facilitate instrument access. The slides are mounted to the aircraft rack and the slide assembly is secured to the aircraft rack with custom retaining brackets and fasteners. For high-altitude aircraft, the angles mount directly to the aircraft rack – photos of the CAFE installation in the left superpod mid-body of the NASA ER-2 are shown in Fig. S3.

### 4.2 Data comparison with ISAF

During the boreal fall deployment (2017) of the NASA Atmospheric TOMography (ATom) campaign (<https://espo.nasa.gov/atom/content/ATom>, last access: 22 August 2019), CAFE joined ISAF on the NASA DC-8 aircraft and sampled from the same inlet. While most ATom flights were in the remote atmosphere where  $[\text{HCHO}]$  is low, the first test flight on 14 September 2017 out of Palmdale, CA, provided good dynamic range with  $[\text{HCHO}]$  ranging from 100 pptv to over 3 ppbv as shown in Fig. 9. CAFE was operated for  $\sim 3$  h of the 4.3 h flight to allow the instrument operator to focus on ISAF, the primary instrument for the campaign. The data in Fig. 9 are plotted as both (a) a time series along with altitude and (b) a cross plot with a linear regression. The time series clearly demonstrates the superior precision of the ISAF measurement. The cross plot and linear regression (York et al., 2004) show a good agreement between the two instruments, with a slope of  $1.04 \pm 0.02$  and a  $y$  intercept of  $24 \pm 7$  pptv.

The agreement was improved through the application of an ozone-dependent correction to the CAFE data, particularly for the time period of 22.5–23 h where the DC-8 passed through stratospherically influenced air. Earlier ATom cir-





**Figure 9.** (a) Time series data from TF01 of ATom-3 show good agreement between CAFE (red) and ISAF (black). (b) A linear regression (York et al., 2004) of the data yields a slope of  $1.04 \pm 0.02$  and intercept of  $0.024 \pm 0.007$ . Data from both instruments were averaged to 10 s.

cuits revealed a strong positive correlation between  $O_3$  and ISAF HCHO in stratospheric air masses, with 0.08 pptv of HCHO produced from exposure to every part per billion by volume of  $O_3$ . Laboratory tests revealed that prolonged exposure to  $O_3$  in the lab eliminated the interference in subsequent flights: no  $O_3$  correction was needed for ISAF during ATom-3. CAFE was not conditioned in a similar manner before ATom-3 and a correction of 1.2 pptv of HCHO per part per billion by volume of  $O_3$  was necessary for ATom-3 data. Possible sources of the higher HCHO yield in CAFE include newer instrument components and the use of THV (a copolymer of tetrafluoroethylene, hexafluoropropylene, and vinylidene fluoride) tubing, which is no longer used in CAFE for sampling.

## 5 Summary

Improvements to the NR-LIF technique for HCHO from St. Clair et al. (2017) include a higher-power laser, fluorescence filters that effectively block  $H_2O$  Raman signal, and a more compact form factor. The new model of the instrument, CAFE, has successfully flown on multiple aircraft platforms and obtained the first NR-LIF airborne intercomparison data with a gold-standard HCHO instrument, ISAF. CAFE currently achieves a  $1\sigma$  precision of 152 pptv for 1 s (87 pptv for 10 s) data at 0 pptv HCHO. The compact size and simple operation make CAFE an ideal instrument for research-grade HCHO measurements from high-altitude aircraft and smaller aircraft focused on local/regional air quality.

**Data availability.** CAFE ATom TF01 data are available at <https://doi.org/10.7910/DVN/DGALD0> (St. Clair, 2019). All other ATom data are available at <https://doi.org/10.3334/ORNLDAAAC/1581> (ATom Science Team, 2017).

**Supplement.** The supplement related to this article is available online at: <https://doi.org/10.5194/amt-12-4581-2019-supplement>.

**Author contributions.** AKS did the mechanical design and assembly of the instrument. SAB wrote the software for the instrument. TFH was responsible for the conceptualization, funding acquisition, and supervision of the project. JMSc wrote the paper as well as the analysis software, and was responsible for the investigation, data curation, and supervision of the project.

**Competing interests.** The authors declare that they have no conflict of interest.

**Acknowledgements.** Jason St. Clair thanks Glenn Wolfe for insightful comments on the paper. Thanks are due to the National Suborbital Research Center for the ATom altitude data and the NOAA UCATS team (James Elkins, Fred Moore, Eric Hintsa) for the  $O_3$  measurements used to correct CAFE ATom data. The aircraft flight opportunity was generously provided by the ATom NASA EVS-2 project.

**Financial support.** This research has been supported by NASA (Upper Atmospheric Research Program grant), NASA (Tropospheric Composition Program grant), and NASA (grant no. NNH12ZDA001N-AITT).

**Review statement.** This paper was edited by Bin Yuan and reviewed by two anonymous referees.

## References

- Aiello, M. and McLaren, R.: Measurement of airborne carbonyls using an automated sampling and analysis system, *Environ. Sci. Technol.*, 43, 8901–8907, <https://doi.org/10.1021/es901892f>, 2009.
- Apel, E. C., Olson, J. R., Crawford, J. H., Hornbrook, R. S., Hills, A. J., Cantrell, C. A., Emmons, L. K., Knapp, D. J., Hall, S., Mauldin III, R. L., Weinheimer, A. J., Fried, A., Blake, D. R., Crounse, J. D., St. Clair, J. M., Wennberg, P. O., Diskin, G. S., Fuelberg, H. E., Wisthaler, A., Mikoviny, T., Brune, W., and Riemer, D. D.: Impact of the deep convection of isoprene and other reactive trace species on radicals and ozone in the upper troposphere, *Atmos. Chem. Phys.*, 12, 1135–1150, <https://doi.org/10.5194/acp-12-1135-2012>, 2012.
- ATom Science Team: ATom NASA DC-8 aircraft data archive, <https://doi.org/10.3334/ORNLDAAC/1581>, 2017.
- Cazorla, M., Wolfe, G. M., Bailey, S. A., Swanson, A. K., Arkinson, H. L., and Hanisco, T. F.: A new airborne laser-induced fluorescence instrument for in situ detection of formaldehyde throughout the troposphere and lower stratosphere, *Atmos. Meas. Tech.*, 8, 541–552, <https://doi.org/10.5194/amt-8-541-2015>, 2015.
- Crawford, J., Davis, D., Olson, J., Chen, G., Liu, S., Gregory, G., Barrick, J., Sachse, G., Sandholm, S., Heikes, B., Singh, H., and Blake, D.: Assessment of upper tropospheric HO<sub>x</sub> sources over the tropical Pacific based on NASA GTE/PEM data: Net effect on HO<sub>x</sub> and other photochemical parameters, *J. Geophys. Res.*, 104, 16255–16273, 1999.
- Fortems-Cheiney, A., Chevallier, F., Pison, I., Bousquet, P., Saunio, M., Szopa, S., Cressot, C., Kurosu, T. P., Chance, K., and Fried, A.: The formaldehyde budget as seen by a global-scale multi-constraint and multi-species inversion system, *Atmos. Chem. Phys.*, 12, 6699–6721, <https://doi.org/10.5194/acp-12-6699-2012>, 2012.
- Fried, A., Walega, J. G., Olson, J. R., Crawford, J. H., Chen, G., Weibring, P., Richter, D., Roller, C., Tittel, F. K., Heikes, B. G., Snow, J. A., Shen, H., O'Sullivan, D. W., Porter, M., Fuelberg, H., Halland, J., and Millet, D. B.: Formaldehyde over North America and the North Atlantic during the summer 2004 INTEX campaign: Methods, observed distributions, and measurement-model comparisons, *J. Geophys. Res.-Atmos.*, 113, 1–16, <https://doi.org/10.1029/2007JD009185>, 2008.
- Fried, A., Barth, M. C., Bela, M., Weibring, P., Richter, D., Walega, J., Li, Y., Pickering, K., Apel, E., Hornbrook, R., Hills, A., Riemer, D. D., Blake, N., Blake, D. R., Schroeder, J. R., Luo, Z. J., Crawford, J. H., Olson, J., Rutledge, S., Betten, D., Biggerstaff, M. I., Diskin, G. S., Sachse, G., Campos, T., Flocke, F., Weinheimer, A., Cantrell, C., Pollack, I., Peischl, J., Froyd, K., Wisthaler, A., Mikoviny, T., and Woods, S.: Convective transport of formaldehyde to the upper troposphere and lower stratosphere and associated scavenging in thunderstorms over the central United States during the 2012 DC3 study, *J. Geophys. Res.-Atmos.*, 121, 7430–7460, <https://doi.org/10.1002/2015JD024477>, 2016.
- Hak, C., Pundt, I., Trick, S., Kern, C., Platt, U., Dommen, J., Ordoñez, C., Prévôt, A. S. H., Junkermann, W., Astorga-Lloréns, C., Larsen, B. R., Mellqvist, J., Strandberg, A., Yu, Y., Galle, B., Kleffmann, J., Lörzer, J. C., Braathen, G. O., and Volkamer, R.: Intercomparison of four different in-situ techniques for ambient formaldehyde measurements in urban air, *Atmos. Chem. Phys.*, 5, 2881–2900, <https://doi.org/10.5194/acp-5-2881-2005>, 2005.
- Hottle, J. R., Huisman, A. J., Digangi, J. P., Kammrath, A., Galloway, M. M., Coens, K. L., and Keutsch, F. N.: Fluorescence-based instrument for in-situ measurements of atmospheric formaldehyde, *Environ. Sci. Technol.*, 43, 790–795, <https://doi.org/10.1021/es801621f>, 2009.
- Jaeglé, L., Jacob, D. J., Brune, W. H., and Wennberg, P. O.: Chemistry of HO<sub>x</sub> radicals in the upper troposphere, *Atmos. Environ.*, 35, 469–489, [https://doi.org/10.1016/S1352-2310\(00\)00376-9](https://doi.org/10.1016/S1352-2310(00)00376-9), 2001.
- Kaiser, J., Li, X., Tillmann, R., Acir, I., Holland, F., Rohrer, F., Wegener, R., and Keutsch, F. N.: Intercomparison of Hantzsch and fiber-laser-induced-fluorescence formaldehyde measurements, *Atmos. Meas. Tech.*, 7, 1571–1580, <https://doi.org/10.5194/amt-7-1571-2014>, 2014.
- Liao, J., Hanisco, T. F., Wolfe, G. M., St. Clair, J., Jimenez, J. L., Campuzano-Jost, P., Nault, B. A., Fried, A., Marais, E. A., Gonzalez Abad, G., Chance, K., Jethva, H. T., Ryerson, T. B., Warneke, C., and Wisthaler, A.: Towards a satellite formaldehyde – in situ hybrid estimate for organic aerosol abundance, *Atmos. Chem. Phys.*, 19, 2765–2785, <https://doi.org/10.5194/acp-19-2765-2019>, 2019.
- Palmer, P. I., Jacob, D. J., Fiore, A. M., Martin, R. V., Chance, K., and Kurosu, T. P.: Mapping isoprene emissions over North America using formaldehyde column observations from space, *J. Geophys. Res.*, 108, 1984–2012, <https://doi.org/10.1029/2002JD002153>, 2003.
- Ren, X., Olson, J. R., Crawford, J. H., Brune, W. H., Mao, J., Long, R. B., Chen, Z., Chen, G., Avery, M. A., Sachse, G. W., Barrick, J. D., Diskin, G. S., Huey, L. G., Fried, A., Cohen, R. C., Heikes, B., Wennberg, P. O., Singh, H. B., Blake, D. R., and Shetter, R. E.: HO<sub>x</sub> chemistry during INTEX-A 2004: Observation, model calculation, and comparison with previous studies, *J. Geophys. Res.*, 113, 1–13, <https://doi.org/10.1029/2007JD009166>, 2008.
- Richter, D., Weibring, P., Walega, J. G., Fried, A., Spuler, S. M., and Taubman, M. S.: Compact highly sensitive multi-species airborne mid-IR spectrometer, *Appl. Phys. B Lasers Opt.*, 119, 119–131, <https://doi.org/10.1007/s00340-015-6038-8>, 2015.
- Shutter, J. D., Allen, N. T., Hanisco, T. F., Wolfe, G. M., St. Clair, J. M., and Keutsch, F. N.: A new laser-based and ultra-portable gas sensor for indoor and outdoor formaldehyde (HCHO) monitoring, *Atmos. Meas. Tech. Discuss.*, <https://doi.org/10.5194/amt-2018-443>, accepted, 2019.
- St. Clair, J. M., Swanson, A. K., Bailey, S. A., Wolfe, G. M., Marrero, J. E., Iraci, L. T., Hagopian, J. G., and Hanisco, T. F.: A new non-resonant laser-induced fluorescence instrument for the airborne in situ measurement of formaldehyde, *Atmos. Meas. Tech.*, 10, 4833–4844, <https://doi.org/10.5194/amt-10-4833-2017>, 2017.
- St. Clair, J.: CAFE instrument HCHO data for TF01 of the NASA ATom-3 aircraft campaign, Harvard Dataverse, <https://doi.org/10.7910/DVN/DGALD0>, 2019.
- Strum, M. and Scheffe, R.: National review of ambient air toxics observations, *J. Air Waste Manage. Assoc.*, 66, 120–133, <https://doi.org/10.1080/10962247.2015.1076538>, 2016.
- Washenfelder, R. A., Attwood, A. R., Flores, J. M., Zarzana, K. J., Rudich, Y., and Brown, S. S.: Broadband cavity-enhanced absorption spectroscopy in the ultraviolet spectral region for mea-

- surements of nitrogen dioxide and formaldehyde, *Atmos. Meas. Tech.*, 9, 41–52, <https://doi.org/10.5194/amt-9-41-2016>, 2016.
- Wolfe, G. M., Nicely, J. M., St. Clair, J. M., Hanisco, T. F., Liao, J., Oman, L. D., Brune, W. H., Miller, D. O., Thames, A., Gonzalez Abad, G., Ryerson, T. B., Thompson, C. R., Peischl, J., McCain, K., Sweeney, C., Wennberg, P. O., Kim, M., Crounse, J. D., Hall, S. R., Ullmann, K., Diskin, G. S., Bui, T. P., Chang, C., and Dean-Day, J.: Mapping hydroxyl variability throughout the global remote troposphere via synthesis of airborne and satellite formaldehyde observations, *P. Natl. Acad. Sci. USA*, 116, 11171–11180, <https://doi.org/10.1073/pnas.1821661116>, 2019.
- York, D., Evensen, N. M., Martínez, M. L., and Delgado, J. D. B.: Unified equations for the slope, intercept, and standard errors of the best straight line, *Am. J. Phys.*, 72, 367–375, <https://doi.org/10.1119/1.1632486>, 2004.